

# Fast Polarimetric Second-Harmonic Generation Microscopy Using Circularly Polarized Light

Kamdin Mirsanaye<sup>1</sup>, Leonardo Uribe Castaño<sup>1</sup>, Ahmad Golaraei<sup>1,2</sup>, Lukas Kontenis<sup>3</sup>,  
Margarete Akens<sup>2</sup>, Susan J. Done<sup>2,4</sup>, Brian Wilson and Virginijus Barzda<sup>1,5</sup>

<sup>1</sup>Department of Physics and Department of Chemical and Physical Sciences, University  
of Toronto, Toronto, Canada

<sup>2</sup>Princess Margaret Cancer Centre, University Health Network, Toronto, Canada

<sup>3</sup>Light Conversion, Vilnius, Lithuania

<sup>4</sup> Department of Medical Biophysics, University of Toronto, Toronto, Canada

<sup>5</sup>Faculty of Physics, Laser Research Centre, Vilnius University, Vilnius, Lithuania

Email: kamdin.mirsanaye@mail.utoronto.ca

**KEY WORDS:** Nonlinear Microscopy, Polarization-Resolved Second-Harmonic Generation, Nonlinear Optical Susceptibility, Biomedical Imaging, Collagen, Muscle.

Polarimetric Second-Harmonic Generation (P-SHG) microscopy enables non-invasive imaging of non-centrosymmetric biological tissues. P-SHG provides detailed ultrastructural information on composition and organization of collagenous and muscular samples, beyond the diffraction limit. The extracted information is often in the form of second order nonlinear susceptibility ratios. This work pertains to the development of a recent P-SHG technique, which greatly enhances imaging speed, by using circularly polarized excitation and emission fields.

The imaging was performed using a high-power amplified laser (Pharos, Light Conversion), coupled to a custom wide-field nonlinear microscope, which enables fast imaging of large fields of view ( $650\mu\text{m} \times 650\mu\text{m}$ ). A unique combination of right and




Figure 1: Polarimetric Second Harmonic Generation Microscopy of Pig Tendon

left handed circular polarization states of the fundamental beam, and the resulting right and left handed circular polarization measurements of the SHG signal allows for the extraction of second order susceptibility ratios, such as  $R = \chi_{zzz}^{(2)} / \chi_{zxx}^{(2)}$  ( $xz$ -plane is the image plane), as well as other quantities such as SHG circular dichroism (SHG-CD) [1]. The analysis of the acquired images is completed in a few seconds, resulting in pixel-resolved R and SHG-CD images, as shown in Fig. 1, that are invaluable in quantitative analysis of biomaterials.

This P-SHG technique allows for fast and precise investigations of ultrastructural variations in the biological tissues, with applications reaching beyond general tissue characterization. Fast imaging of large sample areas is extremely useful in biomedical diagnostics, contractility research, and histopathology [2,3]. In addition, this method may be utilized in whole slide imaging of biological samples, aiding in high throughput quantitative analysis and diagnosis.

[1] L. Kontenis, “Experimental Nonlinear Polarimetric Microscopy,” Dissertation. (2017)

[2] K. Mirsanaye, A. Golaraei, F. Habach, E. Žurauskas, J. Venius, R. Rotomskis, V.

Barzda, “Polar Organization of Collagen in Human Cardiac Tissue Revealed with Polarimetric Second-Harmonic Generation Microscopy”, *Biomed. Opt. Express* **10**(10), 5025–5030. doi: 10.1364/BOE.10.005025 (2019).

[3] Golaraei, et al., “Changes of collagen ultrastructure in breast cancer tissue determined by second-harmonic generation double Stokes-Mueller polarimetric microscopy,” *Biomed. Opt. Express*, **7**(10), 4054. doi:10.1364/boe.7.004054, (2016)