

CIRCULAR-DICHROISM SHG MICROSCOPY PROBES THE SUB-MICROMETER SCALE POLARITY DISTRIBUTION OF OUT-OF-PLANE COLLAGEN FIBRILS.

M. Schmeltz¹, C. Teulon¹, M. Pinsard², U. Hansen³, M. Alnawaiseh⁴, D. Ghoubay⁵, V. Borderie⁵, G. Mosser⁶, C. Aimé⁶, F. Légaré², G. Latour^{1,7}, M.-C. Schanne-Klein¹

(1) LOB, Ecole polytechnique, CNRS, Inserm, IP Paris, 91128 Palaiseau, France

(2) INRS, Centre Energie Matériaux et Télécommunications, Varennes, Qc, Canada

(3) Institute for Musculoskeletal Medicine, Univ. Hospital Münster, Münster, Germany

(4) Department of Ophthalmology, Univ. Münster Medical Center, Münster, Germany

(5) Sorbonne Université, CHNO des Quinze Vingts, INSERM, Institut de la Vision, GRC32, CIC1423, F-75012 Paris, France

(6) Sorbonne Université, CNRS, LCMCP, Paris, F-75005, France

(7) Université Paris-Saclay, 91190 Saint-Aubin, France

E-mail: margauxschmeltz@gmail.com

KEYWORDS: multiphoton microscopy, SHG, collagen, circular dichroism, cornea

Second harmonic generation microscopy (SHG) is now the gold standard technique for in situ visualization of unstained fibrillar collagen in intact tissues [1]. Collagen is the most abundant protein in mammals and is found in many tissues, such as arteries, skin, bone and cornea. The size and the three-dimensional (3D) distribution of collagen fibrils are key distinctive features of every tissue that are crucial for its functional behavior, notably its mechanical properties. In situ characterization of collagen 3D distribution is therefore a major biomedical challenge.

However, usual implementation of SHG microscopy poorly reveals collagen fibrils oriented out of the imaging plane. Recently, circular dichroism SHG (CD-SHG) has been implemented to address this issue [2-5]. This advanced SHG modality measures the normalized difference of the SHG signals excited with left-handed versus right-handed circular polarizations. In this study, a careful implementation of artefact-free CD-SHG imaging of human cornea sections and other collagen-rich samples confirms that it specifically highlights out-of-plane collagen fibrils [5]. We further present a theoretical analysis of CD-SHG signals that goes beyond the electric dipolar approximation to account for collagen chirality and we demonstrate that magnetic dipolar contributions are indeed mandatory to analyze CD-SHG images. We show that the sign of CD-SHG signals does not reveal whether collagen fibrils point upwards or downwards the imaging plane as tentatively proposed previously. CD-SHG instead probes the polarity distribution at the sub-micrometer scale of the out-of-plane fibril assemblies, that is homogeneous polarity versus a mix of anti-parallel fibrils [6]. This makes CD-SHG a powerful tool for characterizing collagen organization in tissues, specifically its degree of disorder. CD-SHG may thus serve to discriminate between healthy and diseased collagen-rich tissues.




Fig. 1 : CD-SHG images of a transverse section of a human cornea in the *up* and *down* configurations. The CD-SHG sign does not change with the out-of-plane orientation.

- [1] S. Bancelin et al, *Nat. Commun.* **5** (2014).
- [2] Chen et al, *Opt. Lett.* **37**, 3837 (2012).
- [2] H. Lee et al, *Biomed Opt Express* **4**, 909 (2013).
- [3] G. Y. Zhuo et al, *Appl. Phys. Lett.* **110** (2017).
- [4] A. Golaraei et al, *Scientific Reports* **9** (2019).
- [5] M. Schmeltz et al, *Opt. Express* **27**, 22685 (2019).
- [6] M. Schmeltz et al, *Optica* **7**, 1469 (2020).